

February 9, 2016

OUR FILE: 46405-9055
Application No: 010313144-001

CMC - Centro Misure Compatibilità Srl
Via della Fisica, 20
Thiene (VI)
Italie
36016

Attention: Renzo Beghetto

Dear Sir/Madame:

The Bureau has received your application for the registration of 3/10m alternative test site. Be advised that the information received was satisfactory to Innovation, Science and Economic Development Canada.

The following number(s) is now associated to the site(s) for which registration / renewal was sought (**Site# 9055A-2**). Please reference the appropriate site number in the body of test reports containing measurements performed on the site. In addition, please keep for your records the following information;

- The company address code associated to the site(s) located at the above address is: **9055A**

Furthermore, to obtain or renew a unique site number, the applicant shall demonstrate that the site has been accredited to ANSI C63.4-2014. A scope of accreditation indicating the accreditation by a recognized accreditation body to ANSI C63.4-2014 shall be accepted. Please indicate in a letter the previous assigned site number if applicable and the type of site (example: 3m OATS or 3m chamber). If the test facility is not accredited to ANSI C63.4-2014, the test facility shall submit test data demonstrating full compliance with ANSI C63.4-2014 standard. The Bureau will evaluate the filing to determine if recognition shall be granted.

The frequency for re-validation of the test site and the information that is required to be filed or retained by the testing party shall comply with the requirements established by the accrediting organization. However, in all cases, test site re-validation shall occur on an interval not to **exceed three years**. There is no fee or form associated with an OATS filing. OATS submissions are encouraged to be submitted electronically to the Bureau using the following URL: http://sms-sgs.ic.gc.ca/eic/site/sms-sgs-prod.nsf/eng/h_00013.html.

If you have any questions, you may contact the Bureau by e-mail at IC.CertificationBureau-Bureauhomologation.IC@ic.gc.ca
Please reference our file and application number above for all correspondence.

Yours sincerely,



Eric Gravel
For: Wireless Laboratory Manager
Certification and Engineering Bureau
3701 Carling Ave., Building 94
P.O. Box 11490, Station AHE
Ottawa, Ontario K2H 8S2
Email: IC.CertificationBureau-Bureauhomologation.IC@ic.gc.ca